NIRT Soybean SRS Bias Log

Tecator Model:_____ Serial # : _____   Location:_____

NOTE: Only results less than two weeks old and within a 5℉ temperature range may be averaged. Results obtained after a slope or bias adjustment, SRS replacement, or NIRT repair or replacement cannot be averaged with results obtained prior to the change (also, the 5-Run correction can only be applied if five consecutive, valid runs are all positive or all negative). Every effort should be made to maintain the bias within the 0.08% Protein and 0.05% Oil tolerances.

Date mm/dd/yy	Temp	RH	Protein Today's Bias $\pm$0.17	Protein 2-Run $\pm$0.12	Protein 3-Run $\pm$0.10	Protein 5-Run $\pm$0.08	S* B* R*	Oil Today's Bias $\pm$0.12	Oil 2-Run $\pm$0.09	Oil 3-Run $\pm$0.07	Oil 5-Run $\pm$0.05	S* B* R*	Oper Init

*(S) slope adjustment, (B) bias adjustment, or (R) SRS replaced. If the NIRT is repaired or replaced, start a new log.